

BED HEAD UNITS


MODELS

ANTEA



Antea Integral design bed unit

Bring efficiency to small spaces




WWW.TEDISELMEDICAL.COM


info@tedisel.com

TECHNICAL DATA

MODEL: ANTEA







- Medical gases
- Electrical outlets (low and high voltage), voice and data connectors
- * General scheme of the Antea model adaptable to the project



Supply unit	
Dimensions	Other lenghts available depending on the project*
Electrical System	
Rated voltage	230 Vca
Rated frequency	50Hz

Medical Gas System	
	O ₂ - Air - N ₂ O - CO ₂ - VAC - AGSS - Air 800 - N ₂ 800 - O ₂ /He - O ₂ /N ₂ O
Materials	
Structure	Aluminium with epoxy painting or anodized (upon request)
Front cover	Aluminium with epoxy painting anodized, HPL or similar material (upon request)



Classification	
Directive 93/42/EEC	CLASS II B
EC Conformity	
Tedisel complies with the provisions of Directive 93/42/EEC (medical devices), ISO 11197 (medical supply units) and IEC 60601-1 (medical electrical equipment. Part1: general requirements for basic safety and essential performance).	

Compatibility with other medical devices	
Tedisel products may be equipped with medical devices from other manufacturers. Please follow the instructions provided by the manufacturers of this equipment for proper installation.	
CE 0197	
TÜV Rheinland LGA Products GmbH	

QUALITY

Tedisel Management System and all our products are certified with quality seal ISO 13485 and CE marking granted by notified body TÜV RHEINLAND LGA Products GmbH.

Tedisel, as medical device manufacturer, has the medical device manufacturer license number 6205-PS, granted by the Spanish Agency for Medicines and Health Products.

All medical devices developed and manufactured in Tedisel, are regulated by the European Directive 93/42/EEC. All equipment has accreditation and certification of electrical safety and electromagnetic compatibility standards ISO 11197 and EN 60601.